

Aplab

OSCILLOSCOPE PROBE (x1/x10 Switchable) 309T



Technical Specifications

x10 POSITION

Bandwidth	: DC to 250MHz.
Input Capacitance	: Nominal 16pF plus Oscilloscope capacitance..
Compensation Range	: 15pF - 35pF.
Attenuation Ratio	: 10:1.
Input Resistance	: 10M ohms (9M ohms probe +1M ohms scope).
Max. Input Range	: 600V (DC + Peak AC).

x1 POSITION

Bandwidth	: DC to 6MHz.
Attenuation	: 1:1.

ADJUSTMENTS

To adjust probe in the x10 position, simply hook the probe at the cal output of your scope and adjust trimmer capacitor at base of connector to best flat waveform.

STANDARD ACCESSORIES

Sprung Hook	: 1 No.
Insulating Tip	: 1 No.
IC Adaptor	: 1 No.
Adjusting Tool	: 1 No.
Ground Spring	: 1 No.

WE PURSUE A POLICY OF CONTINUOUS DEVELOPMENT AND PRODUCT IMPROVEMENT. THUS THE SPECIFICATIONS IN THIS DOCUMENT AND THE LOCATION OF CONTROLS ON THE FRONT PANEL MAY BE CHANGED WITHOUT NOTICE.

Aplab

Test & Measurement Instruments

APLAB LIMITED

APLAB HOUSE, A-5, WAGLE ESTATE, THANE 400 604. INDIA. TEL : 25821861 FAX : 91-022-2582 3137
EMAIL : tmsales@aplab.com WEBSITE : www.aplab.com

- BANGALORE Tel. : 080-23576001 to 07 Fax : 23576008 E-mail : aplabblr@aplab.com
- CHENNAI Tel. : 044-26680472, 26680477 Fax : 26680473 Email : aplabchn@aplab.com
- KOLKATA Tel. : 033-22455435, 22848834 Fax : 22454294 Email : aplabcal@aplab.com
- MUMBAI Tel. : 022-29201787, 29204642, 29203468 Fax : 29209066 Email : aplabwr@aplab.com
- NEW DELHI Tel. : 011-23515183/86, 23627467 Fax : 23634709 Email : aplabnd@aplab.com
- SECUNDERABAD Tel. : 040-27811012, 27843351, 27720339 Fax : 27897788 Email : aplabsec@aplab.com